



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/669,543

09/23/2003

Michael P. Wallace

03-398 (US01)

4647

41696 7590 01/09/2009

VISTA IP LAW GROUP LLP
12930 Saratoga Avenue
Suite D-2
Saratoga, CA 95070

EXAMINER

YABUT, DIANE D

ART UNIT

PAPER NUMBER

3734

MAIL DATE

DELIVERY MODE

01/09/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/669,543	Applicant(s) WALLACE, MICHAEL P.	
	Examiner DIANE YABUT	Art Unit 3734	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 19 May 2008.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-6,10,11,13-16 and 18 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-6,10-11,13-16,18 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

This action is in response to applicant's amendment received on 05/19/2008.

The examiner acknowledges the amendments made to the claims.

Claim Rejections - 35 USC § 103

1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. Claims 1-2 and 6 are rejected under 35 U.S.C. 103(a) as being unpatentable over **Lee** (U.S. Patent No. **6,397,107**) in view of **Engelson** (U.S. Patent No. **5,749,894**) and **Ragheb et al.** (U.S. Patent No. **5,824,049**).

Claims 1-2 and 6: Lee discloses detaching a vaso-occlusive device ("coil") **3** from a delivery catheter **1** to thereby deploy the vaso-occlusive device at a target site in a vasculature of a body including an aneurysm, and delivering energy from an energy emitting element ("induction coil") **13** located outside the body to thereby heat the vaso-occlusive device at the target site (Figures 1-2 and col. 2, lines 50-65).

Lee does not expressly disclose the vaso-occlusive device comprising a therapeutic bioactive agent coating and a polymeric material coating substantially covering the bioactive agent coating, wherein the bioactive agent is released or activated at the treatment site when the vaso-occlusive device is heated or when the polymeric material at least partially melts or softens, wherein the energy-emitting

Art Unit: 3734

element is a radio frequency device. Lee also does not expressly disclose melting and fusing together portions of the coil for stability in a three-dimensional shape.

Engelson teaches a vaso-occlusive device or coil comprising a polymeric material coating which is heated by an energy-emitting element, or a radio frequency device (see abstract; col.1, lines 4-17, col. 4, lines 11 to col. 5, line 64, and col. 6, lines 1-10). The delivered heating energy causes a coating on the vaso-occlusive device to at least partially melt or soften (col. 9, lines 20-37). Engelson also teaches melting and fusing together portions thereof to stabilize the coil in the three-dimensional shape. It would have been obvious to one of ordinary skill in the art at the time of invention to provide a polymeric coating on the vaso-occlusive device heated by radio frequency energy, as taught by Engelson, to Lee since it was well known in the art to use a polymeric material for endovascular therapy that acts as the glue or a means for making the vaso-occlusive device self-adherent without causing substantial harm to the tissue or the blood (col. 4, lines 26-29), and that bioactive agents may be advantageously released by heating the coil.

In addition, it is well known in the art to provide bioactive materials on a device in order to provide advantageous effects on the body, such as promoting healing and preventing infection, and therefore it would be obvious to one of ordinary skill in the art to modify the invention of Lee with bioactive agents. Furthermore, Ragheb et al. teach a device **12** with a therapeutic bioactive agent coating **18** being substantially covered by a polymeric coating **20**. It would have been obvious to one of ordinary skill in the art at the time of invention to cover a bioactive agent coating with a polymeric material

Art Unit: 3734

coating, as taught by Ragheb et al., to Lee to protect the bioactive agent coating from being released until desired, and therefore more effectively delivering the therapeutic agents to the surgical site after being navigated through the vasculature (col. 3, lines 15-19).

3. Claims 3-4, 10-11, and 13 are rejected under 35 U.S.C. 103(a) as being unpatentable over **Lee** (U.S. Patent No. **6,397,107**) in view of **Engelson** (U.S. Patent No. **5,749,894**) and **Ragheb et al.** (U.S. Patent No. **5,824,049**), as applied to claim 1 above, and further in view of **Yamasaki** (U.S. Pub. No. **2004/0215124**).

Claims 3, 10-11, 13: Lee, Engelson, and Ragheb et al. disclose the claimed invention, including sufficiently heating the device to cause coagulation of blood at the target site (col. 2, lines 60-65), except for the energy-emitting element comprising a magnetic resonance device, positioning the body in a magnetic resonance imaging (“MRI”) device, and activating the MRI device to apply a variable magnetic field to the body.

Yamasaki teaches a method and apparatus for aneurismal treatment using a heating source or energy-emitting element comprising a magnetic resonance device (page 7, paragraph 88). It would have been obvious to one of ordinary skill in the art at the time of invention to modify the energy-emitting element outside the body in Lee by providing a magnetic resonance device, as taught by Yamasaki, since it was known in the art that an MRI source yields higher temperature to effectively heat materials that

Art Unit: 3734

are commonly used in vaso-occlusive devices that absorb MRI and radiofrequency energy.

Yamasaki also teaches positioning the body in a magnetic resonance imaging ("MRI") device and activating the MRI device to apply a variable magnetic field to the body, which would thereby heat a highly resistive element in the vaso-occlusive device (page 7, paragraph 88). It would have been obvious to one of ordinary skill in the art at the time of invention to provide a magnetic resonance device as a source, as taught by Yamasaki, to Lee, Engelson, and Ragheb et al. since it was known in the art that an MRI source yields higher temperature to effectively heat materials that are commonly used in vaso-occlusive devices that absorb MRI and radiofrequency energy.

Claim 4: Lee, Engelson, and Ragheb et al. disclose the claimed invention except for the vaso-occlusive device comprising a ferrous material in sufficient concentration to cause heating of the device in response to energy delivered by the magnetic resonance device.

Yamasaki teaches a ferrous material ("magnetically responsive particles Fe_3O_4 ") in sufficient concentration to cause heating of the device in response to energy delivered by the magnetic resonance device (page 1, paragraph 6). It would have been obvious to one of ordinary skill in the art at the time of invention to provide a vaso-occlusion device comprising a ferrous material that responds to applied energy so that the device remains cohesive (page 1, paragraph 5) and therefore properly treating the target site.

Art Unit: 3734

4. Claim 5 is rejected under 35 U.S.C. 103(a) as being unpatentable over **Lee** (U.S. Patent No. **6,397,107**) in view of **Engelson** (U.S. Patent No. **5,749,894**) and **Ragheb et al.** (U.S. Patent No. **5,824,049**), as applied to claims 1 and 19 above, and further in view of **Maitland** (U.S. Patent No. **6,740,094**).

Claim 5: Lee, Engelson, and Ragheb et al. disclose the claimed invention except for the energy-emitting element comprising an ultrasound device acoustically coupled to an exterior of the body.

Maitland teaches a source comprising an ultrasound device acoustically coupled to an exterior of a body, which allows for both local and remote heating (col. 6, lines 26-57). It would have been obvious to one of ordinary skill in the art at the time of invention to modify the energy-emitting element outside the body in Lee by providing an ultrasound device source, as taught by Maitland, in order to allow for both local or remote heating and since it was known in the art that acoustic/ultrasound waves effectively heat materials that are commonly used in vaso-occlusive devices that absorb ultrasonic energy.

Art Unit: 3734

5. Claims 14-16 and 18 are rejected under 35 U.S.C. 103(a) as being unpatentable over **Lee** (U.S. Patent No. **6,397,107**) in view of **Engelson** (U.S. Patent No. **5,749,894**), **Ragheb et al.** (U.S. Patent No. **5,824,049**), and **Ken** (U.S. Patent No. **5,853,418**).

Claims 14-16 and 18: Lee, Engelson, and Ragheb et al. disclose the claimed invention, including detaching a vaso-occlusive device from a delivery catheter to thereby deploy the vaso-occlusive device in an aneurysm and applying magnetic field energy to the device from an energy-emitting element located outside of the body, and a highly conductive metallic coil **3** having a lumen that interacts with a magnetic field to form an eddy current and melting and fusing together portions thereof to stabilize the coil in the three-dimensional shape (see paragraph 2 above) and sufficiently heating the device to cause coagulation of blood at the target site (Lee, col. 2, lines 60-65), but do not disclose a highly resistive element comprising a ferrous material at least partially disposed in the lumen which is heated by way of convective heat transfer from the highly resistive element.

Ken teaches a resistive element **214** containing ferrous material in the lumen of a highly conductive coil **202** which may be heated by way of convective heat transfer from the highly resistive element (Figure 1C; col. 4, line 61 to col. 5, line 10). Although Ken does not expressly disclose the resistive element being of a sufficient concentration to cause heating of the device in response to energy delivered by the magnetic resonance device, it would have been obvious to one of ordinary skill in the art at the time of invention to provide a resistive element within the lumen of the coil in Lee in order for the material to respond to applied energy so that the device remains cohesive

Art Unit: 3734

to properly treat the target site and also since it was well known to provide a concentration of magnetically responsive material high enough so that the device may be heated more rapidly. It is also noted that the "modest amounts of iron" disclosed in Ken does not necessarily render the device to have an insufficient concentration of iron to cause heating in response to magnetic energy.

Response to Arguments

6. Applicant's arguments with respect to claims 1-6, 10-11, 13-16, and 18 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to DIANE YABUT whose telephone number is (571)272-6831. The examiner can normally be reached on M-F: 9AM-4PM EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Todd Manahan can be reached on (571) 272-4713. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3734

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Diane Yabut/
Examiner, Art Unit 3734

/Todd E Manahan/
Supervisory Patent Examiner, Art Unit 3731